Downtime Report for Run 1998-5

October 22, 1998 to December 14, 1998

Scheduled User Time = 1093.6 hours
User Downtime = 106.21 hours
System Downtime = 112.86 hours

Legend:
- Run 1998-5
- Budget

Downtime Budget

1.0% Intentional Refills
0.2% Controls
0.2% Diagnostics
0.1% OAG
0.1% Operations
1.0% Power Supplies
1.0% RF
0.0% Survery & Align.
0.3% Vacuum
0.8% Water/ME
0.1% XFD
0.2% Other

5.0% Total

Y-axis: Percentage of Scheduled User Time (0.0% to 10.0%)

X-axis categories: Scheduled Refills, Controls, Diagnostics, OAG, Operations, PS, RF, Vacuum, Water, Beamline, Other

14-Dec-1998