

Summary of FY2000 Operational Statistics

	Scheduled User Hours	Storage Ring Availability (hours)	Storage Ring Availability (%)	X-ray Availability (hours)	X-ray Availability (%)	Faults	Average Fill Duration without a Fault (hours)	Faults Per Day of Delivered Beam	Delivered Integrated Current (A-hr)	Average Current (ma)
Run 99-5	1024.0	983.0	96.0%	970.8	94.8%	44	22.1	1.09	82.0	84.5
Run 00-1	1511.0	1474.0	97.6%	1461.8	96.7%	35	41.8	0.57	119.5	81.7
Run 00-2	1248.0	1198.7	96.1%	1181.7	94.7%	47	25.1	0.95	99.5	84.2
Run 00-3	1264.2	1161.0	91.8%	1109.5	87.8%	34	32.6	0.74	89.2	80.4
Fiscal Year	5047.2	4816.7	95.4%	4723.8	93.6%	160	29.5	0.81	390.2	82.6