

Summary of FY2013 Operational Statistics

Run Identifier	Scheduled User Hours	Storage Ring Availability [hours]	Storage Ring Availability [%]	X-ray Availability [hours]	X-ray Availability [%]	Injector Availability	Faults	Mean Time to Recovery [hours]	Average Fill Duration without aFault [hours]	Faults Per Day of Delivered Beam	Delivered Integrated Current [A-hr]	Average Current [ma]
Run 12-3	1569.0	1535.1	97.8%	1535.1	97.8%	98.1%	19	1.78	80.8	0.30	153.5	100.0
Run 13-1	1727.0	1707.6	98.9%	1707.6	98.9%	99.1%	10	1.94	170.8	0.14	171.2	100.2
Run 13-2	1704.0	1652.2	97.0%	1652.2	97.0%	92.9%	20	2.59	82.6	0.29	162.8	98.5
Fiscal Year	5000.0	4895.0	97.9%	4895.0	97.9%	98.1%	49	2.14	99.9	0.24	487.4	99.6

Updated

8/23/2013

Data through end of Run 2013-2